

Title:

Calibration Certification**Aerojet Quality Note****E06****Rev #****Date****1****11/13/08**

The calibration supplier listed on the purchase order shall perform all calibrations and is not authorized to subcontract any calibration unless approved by Aerojet. A calibration report shall be furnished to Aerojet for each calibration performed and at a minimum include the following information:

- Identification of the unit being calibrated
- Identification of the calibration source traceable to NIST
- Date of calibration
- Conditions "As Found" and "As Left"
- Notification of out-of-tolerance conditions
- Certification that calibrations were performed in accordance with either ANSI/NCSL Z540, ISO 10012 or ISO/IEC 17025 (latest revisions).
- Identification of measurement uncertainties

Rev	Date	Description of Change	Changed By
1	11/13/08	NCSL was NCSC; ISO/IEC 17025 was Guide 25	M. Sakrison

This document is an integral part of the contract (purchase order) in which referenced. **Applicable Revision:** The revision in effect at the time the purchase order is placed.